

Product: Laboratory Equipment

File Reference: TFR924051_006

Manufacturer: Cole-Parmer Ltd
Beacon Road
Stone, Staffordshire
ST15 0SA
United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Cole-Parmer Infrared Hotplate / stirrers 04807-8x series

The object of the declaration described above is in conformity with the relevant United Kingdom Legislation

Electromagnetic Compatibility Directive 2014/30/EU
Low Voltage Directive 2014/35/EU
RoHS Directive 2011/65/EU (including CDD 2015/863)

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:


EN 61326-1:2013 Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

BS EN 61010-1:2010 +A1:2019 Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements

BS EN 61010-2-010:2014 Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-010: Particular requirements for laboratory equipment for the heating of materials

BS EN 61010-2-051:2015 (Applicable to units with stirring functionality) Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring

Signed for and on behalf of the above manufacturer

Year of Marking 2022
Place of Issue Stone, UK
Authorised Representative Martin Card
Title EMEA Quality and Engineering Manager
Signature 
Date 12 MAY 2022

Catalogue Code	Part Code	Description
04807-84	HP-200-IR-L	Infrared hotplate, 230 V a.c. 50 Hz
04807-85	HP-200-IR-L-120	Infrared hotplate, 120 V a.c. 60 Hz
04807-86	SHP-200-IR-L	Infrared hotplate stirrer, 230 V a.c. 50 Hz
04807-87	SHP-200-IR-L-120	Infrared hotplate stirrer, 120 V a.c. 60 Hz