EU Declaration of Conformity

Product: Electrothermal Laboratory Equipment File Reference: TFR904091_001

Manufacturer: Cole-Parmer Ltd.

Beacon Road

Stone

Staffordshire ST15 OSA

United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Digital Electromantles SHM-250D, RS-200D series

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation

Electromagnetic Compatibility Directive 2014/30/EU

Low Voltage Directive 2014/35/EU
Restriction of Hazardous Substances Directive 2011/65/EU

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

EN 61326:2013 Electrical equipment for the measurement, control and Laboratory use – Part 1:

General requirements

EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and

laboratory use – Part 1: General requirements

EN 61010-2-010:2014 Safety requirements for electrical equipment for measurement, control and

laboratory use – Part 2-010: Particular requirements for laboratory equipment

for the heating of materials

EN 61010-2-051:2015 Safety requirements for electrical equipment for measurement, control and

laboratory use - Part 2-051: Particular requirements for laboratory equipment

for mixing and stirring

Signed for and on behalf of the above manufacturer

Original Year of Marking 2000

Place of Issue Stone, United Kingdom

Authorised Representative Martin Card

Title Senior Operations Manager (Technical)

Signature

Date 26th August 2020

| Catalogue Code | Model | Description |
|-------------------|-------------------|---|
| 36204-13 | - | Omni Controller MKII, 230 V a.c. |
| 36204-15 | SHM-250D-250 | Digital Electromantle 250 ml, 230 V a.c. |
| 36204-17 | SHM-250D-250-115 | Digital Electromantle 250 ml, 110 V a.c. |
| 36204-19 | SHM-250D-500 | Digital Electromantle 500 ml, 230 V a.c. |
| 36204-21 | SHM-250D-500-115 | Digital Electromantle 500 ml, 110 V a.c. |
| 36204-23 | SHM-250D-1000 | Digital Electromantle 1000 ml, 230 V a.c. |
| 36204-25 | SHM-250D-1000-115 | Digital Electromantle 1000 ml, 110 V a.c. |
| 36630-01 | RS-200D | STEM Omni Reaction Station, 10 position, 230 V a.c. |
| 36630-02 | RS-200D-115 | STEM Omni Reaction Station, 10 position, 110 V a.c. |

File Reference: TFR904091_001