

Product: Laboratory Equipment

File Reference: TFR926111_001

Manufacturer: Cole Parmer Limited
Beacon Road
Stone
Staffordshire ST15 0SA
United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration **Incubators**
INC-200D series


The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation

Electromagnetic Compatibility Directive	2014/30/EU
Low Voltage Directive	2014/35/EU
RoHS Directive	2011/65/EU (including CDD 2015/863)

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

BS EN 61010-1:2010 +A1:2019	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements
BS EN 61010-2-010:2014	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-010: Particular requirements for laboratory equipment for the heating of materials
BS EN 61010-2-051:2015	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring
BS EN 61326-1:2013	Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

Signed for and on behalf of the above manufacturer

Original Year of Marking	2003
Place of Issue	Stone, UK
Authorised Representative	Martin Card
Title	EMEA Quality and Engineering Manager
Signature	
Date	6 FEB 2023

Catalogue Code	Model	Description
39065-10	INC-200D-M-120	Microplate Incubator, 120 V a.c. 60 Hz
39065-15	INC-200D-M	Microplate Incubator, 230 V a.c. 50 Hz
39454-00	INC-200D-H-120	Hybridisation Oven 120 V a.c. 60 Hz
39454-05	INC-200D-H	Hybridisation Oven 230 V a.c. 50 Hz