Cole-Parmer®

EU Declaration of Conformity

Product Laboratory Equipment File Number TFR926071 001

Manufacturer Cole-Parmer Ltd

Beacon Road

Stone, Staffordshire

ST15 OSA

United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Hybridisation Ovens

(reference the attached list of catalogue numbers)

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation:

Low Voltage Directive2014/35/EUEMC Directive2004/108/ECRoHS Directive2011/65/EC

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

| IEC/EN 61010-1:2010 | Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements. |
|-------------------------|---|
| IEC/EN 61010-2-010:2003 | Particular requirements for laboratory equipment for the heating of materials. |
| IEC/EN 61010-2-051:2003 | Particular requirements for laboratory equipment for mixing and stirring. |
| IEC/EN 61326-1:2006 | Electrical equipment for measurement, control and laboratory use. EMC requirements. Part 1: General requirements (Class A). |

Signed for and on behalf of the above manufacturer

Additional Information Year of CE Marking: 1996

Place of Issue Stone, Staffordshire, UK

Date of Issue October 2009. Revised 29 September 2017

Authorised Representative Carl Warren

Title Technical Manager

Signature

| Catalogue Number | Model | Description |
|---------------------|------------------|--|
| 39455-01 | HI-200D-HC-T-120 | Hybridiser Hybridisation Oven with Tubes, 120V |
| 39455-04 | HI-200 | Hybrigene Hybridisation Oven only, 230V |
| 39455-06 | НІ-200-Т | Hybrigene Hybridisation Oven with Tubes, 230V |
| 39455-07 | HI-200-120 | Hybrigene Hybridisation Oven only, 120V |
| 39455-08 | HI-200-T-120 | Hybrigene Hybridisation Oven with Tubes, 120V |