


# Cole-Parmer®

## HI-200 Series

### Hybridisation Incubator



Instruction Manual  
Version 8.2

**Cole-Parmer®**  
essentials

## CONTENTS

### Contents

<b>INTRODUCTION.....</b>	<b>4</b>
<b>BEFORE USE .....</b>	<b>5</b>
UNPACKING .....	5
<b>SAFETY AND INSTALLATION .....</b>	<b>6</b>
<b>WARNING .....</b>	<b>6</b>
<b>OPERATOR SAFETY .....</b>	<b>6</b>
<b>INSTALLATION.....</b>	<b>6</b>
Replacement cable.....	7
<b>ENVIRONMENTAL CONDITIONS .....</b>	<b>7</b>
<b>SÉCURITÉ ET CONSIGNES D'INSTALLATION.....</b>	<b>8</b>
<b>AVERTISSEMENT .....</b>	<b>8</b>
<b>INSTALLATION.....</b>	<b>8</b>
Câble de rechange .....	9
<b>CONDITIONS ENVIRONNEMENTALES .....</b>	<b>9</b>
<b>SICHERHEITS- UND INSTALLATIONSFREMDEN INFORMATIONEN.....</b>	<b>10</b>
<b>ACHTUNG .....</b>	<b>10</b>
<b>SICHERHEIT DES BEDIENPERSONALS .....</b>	<b>10</b>
<b>INBETRIEBAHME .....</b>	<b>10</b>
Ersatzkabel.....	11
<b>UMWELTBEDINGUNGEN .....</b>	<b>11</b>
<b>INFORMAZIONI SULLA SICUREZZA E L'INSTALLAZIONE.....</b>	<b>12</b>
<b>AVVERTENZA.....</b>	<b>12</b>
<b>SICUREZZA DELL'OPERATORE.....</b>	<b>12</b>
<b>INSTALLAZIONE .....</b>	<b>12</b>
Cavo di ricambio .....	13
<b>CONDIZIONI AMBIENTALI.....</b>	<b>13</b>
<b>INFORMACIÓN DE SEGURIDAD E INSTALACIÓN.....</b>	<b>14</b>
<b>ADVERTENCIA.....</b>	<b>14</b>
<b>SEGURIDAD DEL OPERARIO .....</b>	<b>14</b>
<b>INSTALACIÓN .....</b>	<b>14</b>
Cable de repuesto .....	15
<b>CONDICIONES AMBIENTALES .....</b>	<b>15</b>
<b>SPECIFICATIONS .....</b>	<b>16</b>
<b>INSTALLATION .....</b>	<b>17</b>
See also the sections on Safety and Installation.....	17
<b>FITTING STACKING PLATES AND WALL MOUNTING PLATES .....</b>	<b>17</b>
Fixing Procedure for two Hybrigenes .....	18
Fixing Procedure for three Hybrigenes .....	19
<b>OPERATION.....</b>	<b>19</b>
<b>THE FRONT PANEL CONTROLS .....</b>	<b>19</b>
<b>SETTING THE TEMPERATURE .....</b>	<b>20</b>
<b>SETTING THE SPEED .....</b>	<b>20</b>
<b>HYBRIDISATION TUBES .....</b>	<b>21</b>
Large (80mm diameter) tube assembly (FHB16) .....	21
Small (44mm diameter) tube assembly (FHB36) .....	21
Mini (4 x 32mm) tube assembly (FHB41 and FHB61).....	22
16 x 15ml tube assembly (F15ML4TH) and 8 x 50ml tube assembly (F50ML4TH) .....	22
<b>INSTALLING THE HYBRIDISATION TUBES .....</b>	<b>23</b>

<b>BASIC PROTOCOL FOR SOUTHERN HYBRIDISATIONS .....</b>	<b>23</b>
Insertion of filters .....	23
Pre-hybridisation.....	23
Hybridisation.....	23
Washes.....	23
Notes: .....	23
<b>WARNING .....</b>	<b>23</b>
<b>USER MAINTENANCE .....</b>	<b>24</b>
O ring seals on the FHB16 hybridisation tubes .....	24
Hybridisation tubes .....	24
Cleaning your Hybridisation Oven.....	24
If you spill the fluid .....	24
<b>ADDITIONAL INFORMATION .....</b>	<b>25</b>
<b>FAULT FINDING .....</b>	<b>25</b>
<b>FUSES .....</b>	<b>25</b>
<b>SAFETY OVER-TEMPERATURE CUT-OUT .....</b>	<b>25</b>
<b>SUPPORT CONTACTS.....</b>	<b>25</b>
<b>SPARES AND ACCESSORIES.....</b>	<b>26</b>

## INTRODUCTION

Please read all the information in this manual before using the unit.

The Cole-Parmer HI-200 Series Hybridisation Oven is ideal for blotting techniques in which RNA, DNA or protein are immobilised onto nylon or nitrocellulose filters. Several filters can be hybridised simultaneously in a single tube while the rotating design ensures volumes as low as 5ml can be used and recovered.

The HI-200 Series Hybridisation Oven has unique "slot-in" tubes with adjustable tube rotation speeds of 0 to 20 rpm; controllable to suit your application. The tubes and other accessories can be accommodated together for multiple applications and the unit can hold up to 16 mini or 4 large glass tubes. The temperature can be set from 10°C above ambient to 80°C and offers excellent temperature accuracy and uniformity.

For safety, the unit has a 10mm thick Perspex door to shield the immediate environment from radiation. In addition, protective casing and non-drip tube design minimises risk to the user. Included is a removable drip tray for easy cleaning should spillages occur.

The HI-200 Series Hybridisation Oven has adjustable feet for levelling on uneven surfaces. It is also possible to stack up to three ovens, saving on valuable laboratory space.

The operating instructions in this Operator's Manual cover the following models of the Cole-Parmer HI-200 Series Hybridisation Oven:

Model number	Description	Voltage
HI-200	Hybridisation incubator	230V, 50/60Hz
HI-200-T	Hybridisation incubator, (including 4 large tubes FHB16)	230V, 50/60Hz
HI-200-120	Hybridisation incubator	120V, 50/60Hz
HI-200-T-120	Hybridisation incubator, (including 4 large tubes FHB16)	120V, 50/60Hz

## **BEFORE USE**

Before using the HI-200 Series Hybridisation Oven please make sure you have read this manual carefully. If there is any doubt relating to the proper use of this equipment, the staff at Antylia Scientific Ltd. or your supplier will be happy to assist you.

### **UNPACKING**

When unpacking the unit please ensure that the following have been removed from the packaging:

- HI-200 Series Hybridisation Oven including drip tray
- Mains cables (UK and EU or US)
- Tube of silicon grease
- Hybridisation tubes x 4 (if included)
- Operator's manual
- Decontamination certificate

The user is advised to keep the original packaging in case the instrument ever needs to be returned for service or repair. Antylia Scientific Ltd. accepts no responsibility for damage incurred unless the unit is correctly packed and transported in its original packaging.

## SAFETY AND INSTALLATION

Please read all the information in this manual before using the HI-200 Series Hybridisation Oven.

### WARNING

HIGH TEMPERATURES ARE DANGEROUS: they can cause serious burns to operators and ignite combustible material.

Cole-Parmer have taken great care in the design of these units to protect operators from hazards, but operators should pay attention to the following points:



- USE CARE AND WEAR PROTECTIVE GLOVES TO PROTECT HANDS
- DO NOT use combustible substances near hot objects
- DO NOT operate the instrument in the vicinity of inflammable liquids or gases.
- DO NOT place any liquid directly into the instrument.

### OPERATOR SAFETY

All operators of Cole-Parmer equipment must have available the relevant literature needed to ensure their safety. It is important that only suitably trained personnel operate this equipment, in accordance with the instructions contained in this manual and with general safety standards and procedures. If the equipment is used in a manner not specified by Antylia Scientific Ltd. the protection provided by the equipment to the operator may be impaired.

All Cole-Parmer instruments are designed to conform to international safety requirements and are fitted with an over-temperature cut-out.

If a safety problem should be encountered, switch off the unit at the mains socket and remove the plug from the electricity supply.

### INSTALLATION

The instrument should be carried using both hands. Never move or carry the instrument when in use or connected to the mains electricity supply.

1. All Cole-Parmer instruments are supplied with a power cable; this may be integral or plug-in.
2.  Before connecting the instrument to the mains electricity supply, check the voltage against the rating plate (located on the back of the unit). **Please note that the unit must be earthed to ensure proper electrical safety.** Connect the mains cable to a suitable plug according to the table below.

Connections	230V	120V
Live	Brown	Black
Neutral	Blue	White
Earth	Green/yellow	Green

UK ONLY: The fused plug supplied with the mains cable is fitted with a 5 amp fuse to protect the instrument and the operator.

Note that units marked 230V on the rating plate work at 220V; units marked 120V work at 110V. In both cases, however, the heating rate will degrade by approximately 8%.

3. Place the unit on a suitable flat bench or in a fume cupboard if required, ensuring that the air inlet vents on the underside are free from obstruction.
4. Plug the mains cable into the socket on the back of the instrument.
5. Switch on the instrument. Symbols on or near the power switch of the unit have the following meanings:

- I Mains Switch On
- O Mains Switch Off

#### **Replacement cable**

Should the mains lead need replacement, a cable of 1mm<sup>2</sup> of harmonized code H05VV-F connected to an IEC320 plug should be used. **IF IN DOUBT CONSULT A QUALIFIED ELECTRICIAN.**

#### **ENVIRONMENTAL CONDITIONS**

The HI-200 Series Hybridisation Oven is designed operate under the following conditions:

- Indoor use
- Ambient temperature range +5°C to +40°C
- Altitude to 2000m
- Relative humidity up to 80%
- Mains supply fluctuations not exceeding 10%
- Over voltage category II IEC 60364-4-443
- Pollution degree 2
- Use with a minimum distance all-around of 200mm from walls or other items

**Note:** The control specifications are quoted at an ambient temperature of 20°C. The specification may deteriorate outside an ambient temperature of between 18°C and 30°C.

The instrument has been tested for radio frequency interference and is certified under EN61326.

# SÉCURITÉ ET CONSIGNES D'INSTALLATION

Veuillez lire attentivement toutes les instructions de ce document avant d'utiliser le séries HI-200 Hybridisation Oven.

## AVERTISSEMENT

Les TEMPÉRATURES ÉLEVÉES SONT DANGEREUSES car elles peuvent provoquer de graves brûlures chez l'opérateur et enflammer les matériaux combustibles.

Cole-Parmer a apporté un soin tout particulier à la conception de ces appareils de façon à assurer une protection maximale des opérateurs, mais il est recommandé aux utilisateurs de porter une attention spéciale aux points suivants :



- PROCÉDER AVEC PRUDENCE ET PORTER DES GANTS POUR SE PROTEGER LES MAINS
- NE PAS utiliser de matériaux combustibles auprès d'objets chauds.
- NE PAS utiliser l'appareil à proximité de liquides ou de gaz inflammables
- NE PAS verser de liquides directement dans l'appareil.

## SECURITE DE L'OPÉRATEUR

Tous les utilisateurs de produits Cole-Parmer doivent avoir pris connaissance des consignes et instructions nécessaires à la garantie de leur sécurité. Important: cet appareil doit impérativement être manipulé par un personnel qualifié et être utilisé selon les instructions données dans ce document, en accord avec les normes et procédures de sécurité générales. Dans le cas où cet appareil ne serait pas utilisé selon les consignes précisées par Antylia Scientific Ltd., la protection pour l'utilisateur ne serait alors plus garantie.

Tous les appareils Cole-Parmer sont conçus pour répondre aux normes de sécurité internationales et sont dotés d'un coupe-circuit en cas de surchauffe.

En cas de problème de sécurité, couper l'alimentation électrique au niveau de la prise murale et enlevez la prise connectée à l'appareil.

## INSTALLATION

Porter l'appareil à deux mains. Ne jamais déplacer ou transporter l'appareil lorsqu'il est en fonctionnement ou branché à l'alimentation électrique.

1. Tous les appareils Cole-Parmer sont livrés avec un câble d'alimentation, qui peut être intégré à l'appareil ou à raccorder.
2.  Avant de raccorder l'appareil à l'alimentation électrique sur secteur, vérifier la tension requise indiquée sur la plaque d'identification (située au dos de l'appareil). **Il est important que l'appareil soit relié à la terre pour assurer la protection électrique requise.** Brancher le câble secteur sur une prise appropriée, voir tableau ci-après.

Connexions	230 V	120 V
Phase	Marron	Noir
Neutre	Bleu	Blanc
Terre	Vert/jaune	Vert

ROYAUME-UNI SEULEMENT: La prise avec fusible intégré fournie avec le câble secteur est munie d'un fusible 5 A destiné à protéger l'appareil et l'utilisateur.

Remarque : les appareils dont la plaque indique 230 V peuvent fonctionner sur 220 V, et ceux dont la plaque indique 120 V peuvent fonctionner sur 110 V. Dans les deux cas cependant, la capacité de chauffage diminuera d'environ 8 %.

3. Placer l'appareil sur une surface plane ou si nécessaire sous une hotte d'aspiration, veiller à ce que les trous d'aération situés sous l'appareil ne soient pas obstrués.
4. Raccorder le câble d'alimentation à la prise située à l'arrière de l'appareil.
5. Allumer l'appareil. Les symboles situés sur ou à côté de l'interrupteur de l'appareil ont la signification suivante :

- |   |   |
|---|---|
| I | Interrupteur secteur en position Marche |
| O | Interrupteur secteur en position Arrêt  |

#### **Câble de rechange**

S'il s'avère nécessaire de remplacer le cordon d'alimentation, utiliser un câble de 1 mm<sup>2</sup> conforme à la norme H05VV-F relié à une prise IEC320. **EN CAS DE DOUTE, CONSULTER UN ELECTRICIEN QUALIFIÉ.**

#### **CONDITIONS ENVIRONNEMENTALES**

La série HI-200 Hybridisation Oven est conçu pour fonctionner dans les conditions suivantes:

- Pour un usage intérieur seulement
- Température ambiante +5°C à +40°C
- Altitude inférieure à 2000m
- Humidité relative ne dépassant pas 80%
- Fluctuations de l'alimentation n'excédant pas 10% de la valeur nominale
- Catégorie II IEC 60364-4-443 de surtension
- Degré de pollution 2
- Utiliser à une distance minimale sur le pourtour de 200mm par rapport aux murs et autres équipements

**Remarque:** Les paramètres sont indiqués pour une température ambiante de 20°C. Ces caractéristiques peuvent se détériorer en dehors d'une température ambiante de 18 à 30°C.

L'appareil a été testé en matière de radiofréquences et est certifié selon la norme EN61326.

# SICHERHEITS- UND INSTALLATIONSGEWINDE

Lesen Sie diese Anleitung vor Verwendung des Serie HI-200 Hybridisation Oven bitte sorgfältig durch.

## ACHTUNG

HOHE TEMPERATUREN STELLEN EINE GEFAHRENQUELLE DAR. Sie können schwere Brandverletzung verursachen und brennbare Stoffe entzünden. Cole-Parmer hat bei der Konstruktion dieses Gerätes sehr darauf geachtet, daß der Bediener vor Gefahren geschützt ist. Dennoch sollten Sie auf die folgenden Punkte achten:



- UMSICHTIG VORGEHEN UND SCHUTZHANSCHUHE TRAGEN
- KEINE brennbaren Stoffe in der Nähe heißer Gegenstände verwenden
- Das Gerät NICHT in der Nähe entzündlicher Flüssigkeiten oder Gase betreiben
- Flüssigkeiten NICHT direkt auf das Gerät auftragen.

## SICHERHEIT DES BEDIENPERSONALS

Alle Benutzer von Cole-Parmer Geräten müssen Zugang zu der entsprechenden Literatur haben, um ihre Sicherheit zu gewähren. Es ist wichtig, daß diese Geräte nur von entsprechend geschultem Personal betrieben werden, das die in dieser Gebrauchsanweisung enthaltenen Maßnahmen und allgemeine Sicherheitsbestimmungen und - vorkehrungen beachtet. Wenn das Gerät anders eingesetzt wird als vom Hersteller empfohlen, kann dies die persönliche Sicherheit des Anwenders beeinträchtigen.

Die Geräte von Cole-Parmer entsprechen den internationalen Sicherheitsbestimmungen und sind mit einem automatischen Übertemperaturabschalter ausgestattet.

Wenn ein Sicherheitsproblem auftreten sollte, muß das Gerät ausgeschaltet und vom Stromnetz getrennt werden.

## INBETRIEBNAHME

Das Gerät mit beiden Händen tragen. Das Gerät unter keinen Umständen transportieren, wenn es in Betrieb ist, oder während das Gerät noch am Netz angeschlossen ist.

1. Alle Geräte von Cole-Parmer werden mit einem Netzkabel geliefert, das entweder eingesteckt wird oder fest mit dem Gerät verbunden ist.
2. Vor dem Anschluss bitte kontrollieren, ob die Stromversorgung den Angaben auf dem Typenschild (auf der Geräterückseite) entspricht. **Um die elektrische Sicherheit zu gewährleisten, muss dieses Gerät geerdet werden.** Schließen Sie das Netzkabel entsprechend der folgenden Tabelle an einen geeigneten Stecker an.

Anschluss	230V	120V
Phase	Braun	Schwarz
Neutral	Blau	Weiß
Erde	Grün/Gelb	Grün

NUR FÜR GROSSBRITANNIEN: der mit dem Netzkabel gelieferte Sicherungsstecker enthält eine 5 Amp. Sicherung zum Schutz des Geräts und des Anwenders.

Geräte, die für 230 Volt ausgelegt sind, können auch bei 220 Volt arbeiten, Geräte für 120 Volt auch bei 100 Volt. In beiden Fällen verringert sich die Aufheizrate um ca. 8%.

3. Stellen Sie das Gerät auf einen geeigneten ebenen Tisch oder in einem Abzugsschrank auf und sorgen Sie dafür, dass die Lufteinlassschlitze auf der Geräteunterseite nicht blockiert sind.
4. Stecken Sie das Netzkabel in die Buchse auf der Geräterückseite ein.
5. Schalten Sie das Gerät ein:

- I Netzschalter Ein
- O Netzschalter Aus

#### **Ersatzkabel**

Bei einem eventuellen Austausch des Netzkabels wird ein Kabel vom Typ H05VV-F mit 1 mm<sup>2</sup> Adernquerschnitt und Europastecker (IEC 320) benötigt. **IM ZWEIFELSFALL EINEN ELEKTROFACHMANN HINZUZIEHEN.**

#### **UMWELTBEDINGUNGEN**

Der Serie HI-200 Hybridisation Oven ist für den Einsatz unter folgenden Bedingungen ausgelegt:

- Gebrauch in Innenräumen
- Umgebungstemperatur zwischen +5°C to +40°C
- Höhe: bis zu 2000 m
- Relative Feuchte nicht über 80%
- Netzspannungsschwankungen nicht über 10%
- Überspannungsklasse 2 IEC 60364-4-443
- Verschmutzungsgrad 2
- Der Mindestabstand zwischen dem Gerät und umgebenden Wänden oder Gegenständen muss 200mm betragen

**Hinweis:** Die Gerätespezifikationen beziehen sich auf eine Umgebungstemperatur von 20°C und können sich außerhalb des Bereichs 18°C bis 30°C verschlechtern.

Das Gerät wurde auf HF-Störeinflüsse geprüft und entspricht den EMV-Bedingungen nach EN61326.

# INFORMAZIONI SULLA SICUREZZA E L'INSTALLAZIONE

Leggere attentamente il presente manuale prima di usare il HI-200 Serie Hybridisation Oven.

## AVVERTENZA

Le ALTE TEMPERATURE SONO PERICOLOSE in quanto possono provocare serie ustioni agli operatori e dare fuoco al materiale combustibile.

La Cole-Parmer ha posto particolare cura nel progettare questo strumento, al fine di proteggere gli operatori da eventuali pericoli, ma gli utilizzatori devono prestare attenzione ai seguenti punti:



- PRESTARE ATTENZIONE ED INDOSSARE GUANTI PROTETTIVI PER LE MANI
- NON usare sostanze combustibili vicino ad oggetti caldi
- NON mettere in funzione lo strumento nei pressi di liquidi o gas infiammabili
- NON collocare alcun tipo di liquido direttamente nello strumento.

## SICUREZZA DELL'OPERATORE

Il personale che utilizza l'apparecchiatura Cole-Parmer deve avere a disposizione la documentazione necessaria al fine di assicurare la loro incolumità. È importante che solo personale adeguatamente addestrato utilizzi questo apparecchio, in conformità alle istruzioni contenute in questo manuale e nel rispetto delle normative e procedure generali di sicurezza. Se l'apparecchio è utilizzato in modo non specificato da Antylia Scientific Ltd., la protezione fornita dall'apparecchiatura all'utilizzatore potrebbe essere a rischio.

Tutte le unità Cole-Parmer sono state progettate in conformità ai requisiti internazionali di sicurezza e sono equipaggiate con un interruttore anti surriscaldamento.

Se si dovesse verificare qualche problema di sicurezza, disconnettere l'apparecchio dalla rete.

## INSTALLAZIONE

Occorre trasportare lo strumento usando entrambe le mani. Non spostare né trasportare lo strumento quando è in funzione o collegato all'alimentazione elettrica di rete.

1. Tutti gli strumenti Cole-Parmer sono forniti con un cavo di alimentazione; può essere integrale o plugin.
2.  Prima di collegare lo strumento all'alimentazione elettrica di rete, controllare la tensione confrontandola con la targhetta riportante i valori nominali (si trova sul retro dell'unità). **Notare che al fine di garantire la corretta sicurezza elettrica, occorre che l'unità sia messa a terra.** Collegare il cavo di rete ad una presa idonea secondo la tabella riportata alla pagina successiva.

Connessione	230V	120V
Sotto tensione	Marrone	Nero
Neutro	Blu	Bianco
Terra	Verde/giallo	Verde

SOLO REGNO UNITO: la spina con fusibile fornita con il cavo di rete è dotata di un fusibile da 5 Amp per proteggere lo strumento e l'utente.

Tenere presente che gli apparecchi riportanti sulla targhetta 230 V funzionano a 220V. Gli apparecchi riportanti 120V funzionano a 110V. Comunque, in entrambi i casi la velocità di riscaldamento diminuirà approssimativamente dell'8%.

3. Collocare l'unità su un banco piano idoneo o in una cappa aspirante se necessario, assicurandosi che gli sfinti delle prese d'aria nella parte inferiore non siano ostruiti.
4. Inserire il cavo di rete nella presa che si trova sul retro dello strumento.
5. Accendere lo strumento:  

I	Interruttore di rete Acceso
O	Interruttore di rete Spento

#### Cavo di ricambio

Qualora occorra sostituire il cavo di rete, si dovrà utilizzare un cavo di 1mm<sup>2</sup> codice armonizzato H05VV-F collegato ad una spina IEC 320. **IN CASO DI DUBBIO, RIVOLGERSI A UN ELETTRICISTA QUALIFICATO.**

#### CONDIZIONI AMBIENTALI

Il HI-200 Series Hybridisation Oven è stato progettato per funzionare nelle seguenti condizioni:

- uso interno
- range di temperatura ambiente da +5°C a +40°C
- altitudine massima 2000 m.
- umidità relativa non superiore all'80%
- oscillazione dell'alimentazione di rete non superiore al 10%
- categoria di sovratensione II IEC 60364-4-443
- grado di inquinamento 2
- Con distanza minima tra i lati dell'apparecchio e il muro o altri oggetti di almeno 200mm

**Nota:** le specifiche di controllo sono indicate ad una temperatura ambiente di 20°C. Le specifiche potrebbero peggiorare fuori da una temperatura ambiente compresa tra 18°C e 30°C.

Lo strumento è stato collaudato per interferenze da radiofrequenze ed è certificato secondo la norma EN61326.

# INFORMACIÓN DE SEGURIDAD E INSTALACIÓN

Lea atentamente este manual antes de utilizar el HI-200 serie Hybridisation Oven.

## ADVERTENCIA

LAS ALTAS TEMPERATURAS SON PELIGROSAS, ya que pueden ocasionar quemaduras graves a los operarios y prender el material combustible. Antyia Scientific ha puesto gran cuidado en el diseño de estos aparatos para proteger al usuario de cualquier peligro; aún así se deberá prestar atención a los siguientes puntos:



- TENGA CUIDADO Y LLEVE GUANTES DE PROTECCIÓN PARA PROTEGERSE LAS MANOS
- NO utilice sustancias combustibles cerca de objetos calientes
- NO utilice el instrumento cerca de líquidos o gases inflamables
- NO coloque un líquido directamente en el instrumento.

## SEGURIDAD DEL OPERARIO

Todos los usuarios de equipos Cole-Parmer deben disponer de la información necesaria para asegurar su seguridad. De acuerdo con las instrucciones contenidas en este manual y con las normas y procedimientos generales de seguridad, es muy importante que sólo personal debidamente capacitado opere estos aparatos. De no ser así, la protección que el equipo le proporciona al usuario puede verse reducida.

Todos los equipos Cole-Parmer han sido diseñados para cumplir con los requisitos internacionales de seguridad y traen incorporados un sistema de desconexión en caso de sobre temperatura.

En caso de que surgiera un problema de seguridad, desconecte el equipo de la red.

## INSTALACIÓN

El instrumento se debe transportar con las dos manos. No mueva ni lleve el instrumento cuando se utilice o esté conectado al suministro eléctrico principal.

1. Todos los instrumentos Cole-Parmer se suministran con un cable de alimentación, que puede ser integrado o 'enchufable'.
2.  Antes de conectar el instrumento al suministro eléctrico, compruebe que el voltaje coincida con el indicado en la placa de régimen (situada en la parte trasera de la unidad). **El instrumento debe disponer de una toma de tierra para garantizar la seguridad eléctrica adecuada.** Conecte el cable de alimentación a un enchufe adecuado según la siguiente tabla.

Conexión	230V	120V
Con corriente	Marrón	Negro
Neutro	Azul	Blanco
Toma de tierra	Verde/amarillo	Verde

SÓLO PARA EL REINO UNIDO: El enchufe suministrado con el cable de alimentación incluye un fusible de 5 amperios para ofrecer protección al instrumento y al usuario.

Asegúrese de que los equipos marcados 230V en la placa indicadora funcionan a 220V y de que los equipos marcados 120V funcionan a 110V. No obstante, en ambos casos la velocidad de calentamiento se verá reducida en un 8% aproximadamente.

3. Sitúe la unidad sobre una mesa plana o en una campana de laboratorio si es necesario, y asegúrese de que los orificios de ventilación situados en la parte inferior no tienen ninguna obstrucción.
4. Conecte el cable de alimentación en el enchufe situado en la parte trasera del instrumento.
5. Encienda el instrumento:
  - Interruptor de alimentación encendido
  - Interruptor de alimentación apagado

#### **Cable de repuesto**

Si es necesario sustituir el cable de alimentación, se debe utilizar un cable de 1mm<sup>2</sup> de código armonizado H05VV, conectado a un enchufe IEC320. **EN CASO DE DUDA, PÓNGASE EN CONTACTO CON UN ELECTRICISTA.**

#### **CONDICIONES AMBIENTALES**

El HI-200 serie Hybridisation Oven está diseñado para utilizarse en las condiciones siguientes:

- Uso en interior
- Intervalo de temperatura ambiente +5°C a +40°C
- Altitud: hasta 2000 m
- Humedad relativa no superior al 80%
- Fluctuaciones del suministro eléctrico no superiores al 10%
- Categoría de sobrevoltaje II IEC 60364-4-443
- Nivel de contaminación 2
- Separado de paredes u otros objetos a una distancia mínima de 200mm

**Nota:** Las especificaciones de control corresponden a una temperatura ambiental de 20°C. Las especificaciones pueden empeorar si se utiliza el instrumento fuera del intervalo de temperatura comprendido entre 18°C y 30°C.

Se han realizado pruebas para comprobar la interferencia de radiofrecuencia del instrumento, el cual cumple la normativa EN61326.

## SPECIFICATIONS

Tube capacity	4 (large and small in any combination) 16 mini tubes (when using the FHB61 mini tube holder) 64 x 15ml tubes (in holder) 32 x 50ml tubes (in holder)
Maximum temperature*	80°C
Minimum temperature	10°C above ambient
Temperature selection	Digital
Adjustable rotation speed	0; 5 to 20 rpm
Adjustable rocking platform (optional)	30 opm recommended
Temperature display	LED
Temperature set point resolution	0.1°C
Temperature sensor	PRT
Temperature stability in the chamber	1.0°C
Temperature stability in the tubes	0.1°C
Temperature uniformity in the chamber#	±1.5°C
Temperature uniformity in the tubes#	±0.5°C
Temperature accuracy	±0.3°C
External dimensions (W x D x H)	355 x 383 x 432mm
Electrical supply	230V, 50/60Hz 120V, 50/60Hz
Power	750W
Net weight	15kg
Shipping weight	21kg

**Note:** The HI-200 Series Hybridisation Ovens include a pre-set over-temperature cut-out.

#Temperature measured in the centre of each tube.


\*Large tubes are fitted with seals suitable for use up to 70°C. For higher temperatures, seals part number 6101488 are required.

For the hybridisation tubes see the list under Spares and Accessories

# INSTALLATION

See also the sections on Safety and Installation.

1. Place the unit on a suitable flat bench or in a fume cupboard if required, ensuring that the air inlet and outlet vents are free from obstruction.
2. Place some water in a hybridisation tube and fit the tube in the HI-200 Series Hybridisation Oven. Adjust the feet on the base of the unit by unscrewing them anticlockwise slightly until the unit is level. Check the level by placing the tube in different positions in the oven.
3. Plug the mains cable into the power socket on the rear of the unit.



*Rear view of the Hybridisation Oven*

## FITTING STACKING PLATES AND WALL MOUNTING PLATES

A trained technician should carry out this work only after reading the full instructions.

There are two kits available for stacking and wall fixing either two or three Hybridisation units:

**FHB4DSTK:** Stacking and wall mounting kit for two units


Description	Part code	Quantity
Wall Bracket	6105214	1
M4 x 8 Pozi Pan Screw	6008652	9
M4 Crinkle Washers	6008639	9
Thumbscrew	6009552	1
Wall Plate	6105213	1
Stacking Plate	6105212	1

**FHB4DKIT:** Stacking and wall mounting kit for three units

Description	Part code	Quantity
Wall Bracket	6105214	1
M4 x 8 Pozi Pan Screw	6008652	15
M4 Crinkle Washers	6008639	15
Thumbscrew	6009552	1
Wall Plate	6105213	1
Stacking Plate	6105212	2

#### Fixing Procedure for two Hybrigenes


1. Ensure the units are turned off and remove the power to them.
2. Place one unit on the top of the other.
3. The stacking plate has three holes and three slots. It has three slots to allow for the adjustment of the feet for levelling the units.
4. Use three of the screws to attach the plate to the bottom of the top unit.
5. Use another three screws to attach the plate to the bottom hybridiser.
6. Ensure that the units are level and the screws are tight.



*Fixing the stacking plate*

7. For safety it is recommended that the top unit is fixed to a wall to prevent toppling. Use the final three screws to attach the wall bracket to the top of the top unit. Ensure that the screws are tight.

8. Put the hybridisers in the position you want to use them.
9. Position the wall plate and mark the position of the holes.
10. Use the appropriate screws and fixings to attach the wall plate to the wall; ensure that the screws are tight.



*Hybridisation Oven with wall bracket attached*

11. Re-position the hybridisers placing the wall bracket over the wall plate.
12. Use the thumb-screw to attach the wall bracket to the wall plate.

#### **Fixing Procedure for three Hybrigene**

If you have three units, repeat the above process attaching a stacking plate between each pair of units.


## **OPERATION**

### **SWITCHING ON**

Once your Hybridisation Oven is connected to a power supply, switch it on using the power switch next to the power inlet socket on the rear panel.


### **THE FRONT PANEL CONTROLS**

The front panel controls consist of a speed control knob for adjusting the tube rotation speed and a temperature controller.



*The front panel controls*

## SETTING THE TEMPERATURE



1. Press the 'P' button on the controller and the set temperature will be displayed.
2. To increase the set temperature, press the Increase button.
3. To decrease the set temperature, press the Decrease button.
4. Press P and the unit will now go to the set temperature.
5. If no key is pressed for 5 seconds the display will return to show the actual temperature and the unit will go to the new set temperature.
6. The Heater indicator will be lit while the heater is on.
7. The set temperature can be checked at any time by pressing the 'P' button.
8. Press the Increase and Decrease buttons together to toggle between °F and °C.

**Note:** When the unit is tuned off, on subsequent power-up the units will revert to °C.

## SETTING THE SPEED



*Speed control knob*


1. When turned fully anticlockwise, the speed control knob switches the rotation of the tubes off.
2. To increase the rotation speed, turn the knob clockwise until you reach the required speed.
3. Operation of the speed control knob increases the speed clockwise and decreases the speed anticlockwise in steps of approximately 4 rpm.

## HYBRIDISATION TUBES

The hybridisation tubes are manufactured from durable thick borosilicate glass with a non-drip design. The tubes must be checked regularly and replaced if any damage occurs. Any damage to the ends of the tubes can cause them to break or leak when the end caps are inserted.

### Large (80mm diameter) tube assembly (FHB16)


Remove the tube end cap which has the two holes in it. Place the filters in the tube and replace the end cap. A small quantity of Cole-Parmer O-ring grease (part code 6101351, supplied with the Hybridisation Oven) **must** be smeared on the seal to facilitate the insertion of the end caps. Remove the spindle cap and add the fluids appropriate to the protocol through the hole in the spindle using a syringe. Replace the spindle cap and press it onto the spindle so that it clicks into place. Note: the tube will not rotate properly if the cap is not fitted securely.



*FHB16 assembly*

### Small (44mm diameter) tube assembly (FHB36)


Unscrew the tube end cap. Place the filters in the tube and replace the end cap. Remove the spindle cap and add the fluids appropriate to the protocol through the hole in the spindle using a syringe. Replace the spindle cap and press it onto the spindle so that it clicks into place.



*FHB36 assembly*

### **Mini (4 x 32mm) tube assembly (FHB41 and FHB61)**


The mini tubes must be used in the FHB61 multi-tube holder. Unscrew the end cap of the tube. Place the filters and the liquids appropriate to the protocol in the tube and replace the end cap. Fix the tube into a multi-tube holder, using the springs, and ensure that there are four tubes in the holder. The end caps must be as shown, on alternate ends. The multi-tube holder will not rotate properly if there are not four mini tubes in it.



*4 x FHB41 and FHB61 assembly*

### **16 x 15ml tube assembly (F15ML4TH) and 8 x 50ml tube assembly (F50ML4TH)**

These tube holders are designed to hold 15ml and 50ml Falcon-type tubes. Unscrew the end cap from the tube. Place the filters and the liquids appropriate to the protocol in the tube and replace the end cap. Assemble the tubes into the multi-tube holder using the spring clips to hold them in place. The multi-tube holder will not rotate properly if it is not balanced properly. The multi-tube holder should be filled completely or at least balanced even if all the tubes are not being used.



*F15ML4TH assembly*

## INSTALLING THE HYBRIDISATION TUBES

1. Pre-heat the oven to the required temperature.
2. Insert the tubes or tube holder into the Hybridisation Oven ensuring that they rotate freely and the end caps engage on the drive rollers on the right hand side of the chamber.
3. Close the door.
4. Turn the speed control knob until you reach the required speed.

## BASIC PROTOCOL FOR SOUTHERN HYBRIDISATIONS

### Insertion of filters

Place a wet or moist filter, DNA side facing inwards, into a hybridisation tube and roll the tube, allowing the filter to adhere to the glass. Large trapped air bubbles should be smoothed out with a glass pipette; smaller air bubbles are acceptable and do not interfere with hybridisation so long as the surface of the filter is covered with hybridisation fluid.

Several filters may be placed in one tube; there is a slight loss of signal, but no increase in background in the lower filters. To improve the signal we recommend the use of separation membranes between filters, part number FMEM2020.

### Pre-hybridisation

Add 5 or 10ml of hybridisation fluid, according to the number of filters and the size of tube, so that the filters are covered. Allow to pre-hybridise for the required period.

### Hybridisation

Drain off the pre-hybridisation fluid by removing the spindle cap from the fill and drain hole in the blue end cap and inverting the tube. Add the hybridisation liquid together with the radioactive probe using a syringe and needle through the fill and drain hole and replace the spindle cap. The minimum volume for a single filter is 5ml (when using large tubes). Allow to hybridise for the required period.

### Washes

Drain off the hybridisation fluid through the fill and drain hole, collecting the probe in a suitable container for re-use or disposal. Washing depends on the stringency required and usually involves a series of washes with SSC buffer and SDS of decreasing concentration until background radiation has been removed.

### Notes:

- All radioactive waste can be collected for disposal; none need be disposed of down the sink.
- Radioactive probes can be collected and reused several times.
- As the protocols for this unit usually involve the use of radioactive materials the appropriate manufacturer's safety precautions must be taken into consideration.

## WARNING



THE UNIT WORKS UP TO 80°C. This temperature can cause burns. Antylia Scientific Ltd. takes great efforts to protect operators from hazards, but the following points are brought to your attention:

1. Use care and wear protective gloves to remove hybridisation tubes from the cabinet if the unit is at high temperature. If the protocol allows, it is better to switch off the unit and open the door to allow the unit to cool to below 40°C before attempting to remove the tubes.
2. The coloured end cap of the 80mm large tube and the 44mm small tube has two holes. Take care not to spill radioactive probe through these holes. One hole allows the air to expand when the tube is heated. The other is the fill and drain hole which is revealed when the spindle cap is removed.

**DO NOT TIP THE TUBE SO THAT THE COLOURED END IS DOWN AS THIS WILL CAUSE A SPILLAGE.**

3. The unit has been designed to conform to international safety specifications and is fitted with an over-temperature cut-out. If, for any reason, the temperature rises above a safe level for the instrument, the over temperature protection cuts out the heater, but the tubes continue to rotate. Check the reason for the cut out and, if there is no obvious cause, turn the unit off, allow it to cool then turn it on again. The cut-out should reset itself. If it does not re-set then the unit needs repair.
4. If the Controller displays 'Err', the thermistor has broken or become disconnected and the unit will cool down automatically.

## **USER MAINTENANCE**

### **O ring seals on the FHB16 hybridisation tubes**

The O ring seals on the large 80mm diameter hybridisation tubes must be checked regularly and replaced before a leak occurs. Each tube is provided with 6 spare O rings and further O rings can be obtained from Antylia Scientific Ltd. or your Cole-Parmer dealer.

### **Hybridisation tubes**

The hybridisation tubes must be checked regularly and replaced if any damage occurs. Any damage to the ends of the tubes can cause them to leak or break when the end caps are inserted. Further tubes can be obtained from Antylia Scientific Ltd. or your Cole-Parmer dealer.

### **Cleaning your Hybridisation Oven**

Before cleaning your unit ALWAYS disconnect from the power supply and allow to cool.

You can clean the case of the Hybridisation Oven with a cloth dampened in soapy water. No part of the unit should be immersed in solvents. Do not use acetone or abrasive cleaners.

**DO NOT USE ETHANOL OR METHYLATED SPIRITS ON THE ACRYLIC DOOR;** these may make the acrylic go cloudy.

Before using any cleaning or decontamination method except those recommended in this manual, the responsible body should check with Antylia Scientific Ltd. that the proposed method will not damage the equipment.

### **If you spill the fluid**

Because of the nature of operation of the Hybridisation Oven, the complete inside must be cleaned if a spillage does occur. This is to remove any airborne particles which may have been blown around the system and to guarantee full operator safety.

Switch the Hybridisation Oven off and disconnect it from the power supply. Follow this sequence of operations when the Hybridisation Oven has been allowed to cool below at least 50°C:

- Remove all glassware from the Hybridisation Oven.
- Remove the plastic drip tray.
- Clean all surfaces of the inner chamber using a suitable cleaning and decontaminating solution. Be careful of the drive shafts.

Cleaning and decontamination materials recommended by Antylia Scientific include:

- Lipsol®
- Decon® 90
- Neutracon®
- Other mild detergent

## **ADDITIONAL INFORMATION**

Note that this equipment should only be dismantled by properly trained personnel. Removing the outer cover exposes potentially lethal mains voltages. There are no user serviceable parts within this equipment.

### **FAULT FINDING**

Should you have any problems with your HI-200 Series Hybridisation Oven which cannot be easily remedied, you should contact your supplier and return the unit if necessary. Contact the Antylia Scientific Ltd. Service Department (details given below) to obtain information on how to return the unit. Fill in the Decontamination Certificate provided with the unit and include the serial number and details of the fault observed. Please remember to return the unit in its original packing. Antylia Scientific Ltd. accepts no responsibility for damage to units which are not properly packed for shipping: if in doubt, contact your supplier, giving the full serial number of the unit.

### **FUSES**

If the controller fails to light when power is supplied to the unit the fuses may have blown. Check that there is no external cause such as a faulty plug; check both fuses and replace any faulty fuse with one of the correct value. Fuse values are given on the label next to the mains input socket.

### **SAFETY OVER-TEMPERATURE CUT-OUT**

In the event of no heater power, check the mains plug and lead, then reset the cut-out control by switching the unit off, allow to cool then turn it on again. The cut-out should reset itself. If it does not re-set then the unit needs repair.

Repeated operation of the cut-out indicates a serious fault: you may need to return your unit to your supplier for repair.

### **SUPPORT CONTACTS**

If you require further technical or application assistance please contact Antylia Scientific Ltd at:

E-mail: [cptechsupport@antylia.com](mailto:cptechsupport@antylia.com).

Phone: +44 (0)1785 810433

For servicing information please contact:

E-mail: [cpservice@antylia.com](mailto:cpservice@antylia.com). Phone: +44 (0)1785 810475

Antylia Scientific Ltd.  
Service Department  
Beacon Road  
Stone  
Staffordshire  
ST15 0SA

Please note that neither Antylia Scientific Ltd. nor any of their agents can accept the return of ANY GLASSWARE from these units.

If the unit is returned it must be thoroughly decontaminated and a Decontamination Certificate returned with the unit. A copy of the certificate is available from our website, [www.coleparmer.com](http://www.coleparmer.com), or can be requested from the Service Department if the unit is being returned.

## SPARES AND ACCESSORIES

The following spares and accessories may be purchased from Antylia Scientific Ltd. or their dealers. All Cole-Parmer tubes are made from durable Borosilicate glass. Contact Antylia Scientific Ltd. or your Cole-Parmer dealer for further details.

<b>Part code</b>	<b>Description</b>
FHB16	Large (80mm diameter) hybridisation tube assembly complete with sealing rings and end caps.
FHB15	Large (80mm diameter) glass tube only
6104412	End cap (fill and drain) for large tubes, includes spindle cap (no sealing ring)
6104413	End cap (blank) for large tubes (no sealing ring)
6101387	Spindle cap (for use with 6104412 and 6101486)
6101090	Sealing ring for large tubes (for 70°C operation)
6101488	Sealing ring for large tubes (for >70°C operation)
6101351	O-ring grease (100g tube)
FHB36	Small (44mm diameter) hybridization tube assembly complete with sealing ring and end cap.
FHB35	Small (44mm diameter) glass tube only
6101486	End cap (fill and drain) for small tubes, includes spindle cap (no sealing ring)
6007625	Sealing ring for small tubes
FHB41	4 x 32mm diameter mini tubes
6102395/1	Mini tube, glass tube only
6102375	Cap for use with the mini tube
6102376	Seal for use with the mini tube
FHB61	Multi-tube holder for use with FHB41 (holds 4 mini tubes)
F15ML4TH	16 x 15ml tube assembly (holds 15ml Falcon-type disposable tubes)
F50ML4TH	8 x 50ml tube assembly (holds 50ml Falcon-type disposable tubes)
FHB4/PLAT	Rocking platform
FHB4SH1	Stainless steel mesh shelf
FHTRACK	Tube rack/holder (holds 3 large and 3 small tubes)
FMEM2020	Membrane separators 20 x 20cm (pack of 5)
FHB4DSTK	Stacking and wall mounting kit for two Hybrigenes
FHB4DKIT	Stacking and wall mounting kit for three Hybrigenes
6105112	Drip tray
HH179(S)	UK mains lead with plug, 230V
HH180(S)	European mains lead with plug, 230V
FCABLEUS	US mains lead with plug, 120V



This product meets the applicable CE Directives and UKCA Legislation for radio frequency interference and may be expected not to interfere with, or be affected by, other equipment with similar qualifications. We cannot be sure that other equipment used in its vicinity will meet these standards and so we cannot guarantee

that interference will not occur in practise. Where there is a possibility that injury, damage or loss might occur if equipment malfunctions due to radio frequency interference, or for general advise before use, contact the manufacturer.

Declaration of Conformity is available to view online at [www.coleparmer.com](http://www.coleparmer.com)

**EU Representative address**  
Antylia Scientific GmbH  
Futtererstraße 16  
97877 Wertheim  
Deutschland  
Tel: +49 9377 9203-0  
Email: sales@coleparmer.de

**UK Representative address**  
Antylia Scientific  
9 Orion Court  
Ambuscade Road  
Colmworth Business Park  
St. Neots  
PE19 8YX  
United Kingdom  
Tel: +44 (0) 1480 277339  
Email: [enquiries@antylia.com](mailto:enquiries@antylia.com)

## Ordering Information

Order No.	Series	Model	Legacy SKU
39455-04	HI-200	HI-200	FHB4DD
39455-07	HI-200	HI-200-120	FHB4DP
39455-06	HI-200	HI-200-T	FHB4DDT
39455-08	HI-200	HI-200-T-120	FHB4DPT

### Warranty Registration



**Cole-Parmer®**  
essentials

**Antylia Scientific Ltd**

Beacon Road,  
Stone,  
Staffordshire,  
ST15 0SA,  
United Kingdom

### UK

T: +44 (0) 1480 272279  
E: uk.sales@antylia.com  
W: coleparmer.co.uk

### India

T: +9122 61394444  
E: info@coleparmer.in  
W: coleparmer.in

### Germany

T: +49 (0) 9377 92030  
E: de.sales@antylia.com  
W: coleparmer.de

### China

T: +1 847 549 7600  
E: sales@antylia.com  
W: coleparmer.com

### France

T: +33 (0) 1486 37800  
E: fr.sales@antylia.com  
W: coleparmer.fr

### USA

T: +1 847 549 7600  
E: sales@antylia.com  
W: coleparmer.com

### Italy

T: +39 (0) 284349215  
E: it.sales@antylia.com  
W: coleparmer.it

### Canada

T: +514 355 6100  
E: info@antylia.ca  
W: coleparmer.ca

### Other

T: +1 847 549 7600

