



Declaration of Conformity

Cole-Parmer®

Product: Laboratory Equipment

File Reference: TFR905091_001

Manufacturer: Cole Parmer Limited
Beacon Road
Stone
Staffordshire ST15 0SA
United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration	Relation to GFT series
-----------------------	---------------------------


The object of the declaration described above is in conformity with the relevant United Kingdom Legislation

Electromagnetic Compatibility Regulations	2016 No. 1091
Electrical Equipment (Safety) Regulations	2016 No. 1101
RoHS Regulations	2012 No. 3032 (including 2019 No. 492)

References to the relevant designated standards used or references to the other technical specifications in relation to which conformity is declared:

BS EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements
BS EN 61010-2-051:2015	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-010: Particular requirements for laboratory equipment for mixing and stirring
BS EN 61326-1:2013	Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

Signed for and on behalf of the above manufacturer

Original Year of Marking	2022
Place of Issue	Stone, UK
Authorised Representative	Martin Card
Title	EMEA Quality and Engineering Manager
Signature	
Date	9 NOV 2022

Catalogue Code	Model	Description
98930-40	GT-200-6-120	Gelation timer, 999.9 minutes, 120 V a.c. 60 Hz
98930-41	GT-200-6	Gelation timer, 999.9 minutes, 230 V a.c. 50 Hz
98930-42	GT-200-5-120	Gelation timer, 9999 minutes, 120 V a.c. 60 Hz
98930-43	GT-200-5	Gelation timer, 9999 minutes, 230 V a.c. 50 Hz