

EU Declaration of Conformity



Product: Laboratory Equipment

File Reference: TFR910121_002

Manufacturer: Antylia Scientific Ltd
Beacon Road
Stone, Staffordshire
ST15 0SA
United Kingdom

EU Representative

Antylia Scientific GmbH
Futtererstraße 16
97877 Wertheim
Deutschland

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration

**Cole-Parmer MP-250 series
Melting Point Detectors**

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation

Electromagnetic Compatibility Directive

2014/30/EU

Low Voltage Directive

2014/35/EU

RoHS Directive

2011/65/EU (including CDD 2015/863)

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

EN 61326-1:2013

Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

EN 61010-1:2010
/A1:2019/AC:2019-04

Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements

IEC / EN 61010-2-010:2014

Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-010: Particular requirements for laboratory equipment for the heating of materials

Signed for and on behalf of the above manufacturer

Original Year of Marking

2012

Place of Issue

Stone, UK

Authorised Representative

Martin Card

Title

EMEA Quality and Engineering Manager

Signature

Date

22 Sept 2021

Catalogue Number	Model	Description
03010-15	MP-250D	Mel-Temp melting point detector, fixed ramp rate
03010-51	MP-250D-F	Melting point detector, fixed ramp rate
03010-41	MP-250D-P	Melting point detector, programmable ramp rate
03010-59	MP-250D-PR	Melting point detector, beginning / ending recording