EU Declaration of Conformity



Product: Laboratory Equipment File Reference: TFR904121_001

Manufacturer: Antylia Scientific Ltd

Beacon Road

Stone, Staffordshire

ST15 OSA

United Kingdom

EU Representative Antylia Scientific GmBH

Futtererstraße 16 97877 Wertheim Deutschland

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Cole-Parmer Melting Point
Detectors MP-100 series

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation

Electromagnetic Compatibility Directive 2014/30/EU

Low Voltage Directive 2014/35/EU

RoHS Directive 2011/65/EU (including CDD 2015/863)

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

IEC / EN 61326-1:2013 Electrical equipment for the measurement, control and Laboratory use

- Part 1: General requirements

EN 61010-1:2010/ Safety requirements for electrical equipment for measurement, control and

A1:2019/AC:2019-04 laboratory use – Part 1: General requirements

IEC / EN 61010-2-010:2014 Safety requirements for electrical equipment for measurement, control and

laboratory use - Part 2-010: Particular requirements for laboratory equipment

for the heating of materials

Signed for and on behalf of the above manufacturer

Original Year of Marking 2001

Place of Issue Stone, UK

Authorised Representative Martin Card

Title EMEA Quality and Engineering Manager

Signature

Date 15th March 2021

C	Catalogue Code	Model	Description
	03013-61	MP-100	Analogue Melting Point Apparatus, 230 V a.c.