



Declaration of Conformity

Cole-Parmer®

Product: Laboratory Equipment

File Reference: TFR908051_001

Manufacturer: Cole Parmer Limited
Beacon Road
Stone
Staffordshire ST15 0SA
United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration

**Scanning spectrophotometers
74 and 76 series**

The object of the declaration described above is in conformity with the relevant United Kingdom Legislation

**Electromagnetic Compatibility Regulations
Electrical Equipment (Safety) Regulations
RoHS Regulations**

**2016 No. 1091
2016 No. 1101
2012 No. 3032 (including 2019 No. 492)**

References to the relevant designated standards used or references to the other technical specifications in relation to which conformity is declared:


BS EN 61010-1:2010
+A1:2019

Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements

BS EN 61326-1:2013

Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

Signed for and on behalf of the above manufacturer

Original Year of Marking	2022
Place of Issue	Stone, UK
Authorised Representative	Martin Card
Title	EMEA Quality and Engineering Manager
Signature	
Date	10 OCTOBER 2022

Model	Catalogue Code	Description
741001	83056-21	7410 Vis spectrophotometer, cream
741501	83056-22	7410 UV/Vis spectrophotometer, cream
747501	83056-23	7415 Nano spectrophotometer, cream
741001B	83056-31	7410 Vis spectrophotometer, black
741501B	83056-32	7410 UV/Vis spectrophotometer, black
747501B	83056-33	7415 Nano spectrophotometer, black
761501	83056-26	7615 UV/Vis spectrophotometer, cream
761501B	83056-36	7615 UV/Vis spectrophotometer, black