EU Declaration of Conformity

Product Laboratory Equipment File Number TFR903121 001

Manufacturer Cole-Parmer Ltd

Beacon Road

Stone, Staffordshire

ST15 0SA

United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Flocculator

(reference the attached list of catalogue numbers)

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation:

Low Voltage Directive 2006/95/EC (until 19 April 2016)

2014/35/EU (from 20 April 2016)

EMC Directive 2004/108/EC **RoHS Directive** 2011/65/EC

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

IEC/EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements.
IEC/EN 61010-2-051:2015	Particular requirements for laboratory equipment for mixing and stirring.
IEC/EN 61326-1:2006	Electrical equipment for measurement, control and laboratory use. EMC requirements. Part 1: General requirements (Class A).

Signed for and on behalf of the above manufacturer

Additional Information Year of CE Marking: 2002

Place of Issue Stone, Staffordshire, UK

Date of Issue June 2009. Revised 06 February 2017

Authorised Representative Carl Warren

Title Technical Manager

Signature

Catalogue Number	Model	Description
99551-00	FJT-200D	Flocculation Jar Tester, 2-place, 230 V
99551-05	FJT-200D-6	Flocculation Jar Tester, 6-place, 230 V