

Contour Comparator

Simple to use, and with a wide range of kit options, the Contour Comparator offers an excellent solution to visual testing requirements.

- Compact size with grip feature for ease of handling
- High clarity and durable colour discs
- Choose either a single or dual parameter kit. Each kit contains the instrument disc(s), cuvettes, brush, stirring rods and reagents for 30 tests



Parameter	Measurement Range	Single Parameter Kit (includes 30 tests)	Replacement Discs (each)
Alkalinity	0 - 250 mg/L CaCO ₃	CKH1192	-
Aluminium	0 - 0.5 mg/L Al	CKH1166	CKD1166
Ammonia	0 - 1 mg/L N	CKH1152	CKD1152
Bromine	0-8 mg/L Br ₂	CKD1005	CKD1005
Chlorine - (DPD 1, Free)	0 - 5 mg/L Cl ₂	CKH1001	CKD1001
Chlorine - Free and Total Chlorine	0 - 2 mg/L Cl ₂	CKH1002	CKD1002
Chlorine - Free and Total Chlorine	0 - 1 mg/L Cl ₂	CKH1003	CKD1003
Chlorine HR	0 - 50 mg/L Cl ₂	CKH162/50	CKD162/50
Chlorine HR	0 - 250 mg/L Cl ₂	CKH 162/250	CKD162/250
Copper	0 - 5 mg/L Cu	CKH1186	CKD1186
Fluoride	0 - 1.5 mg/L F	CKH1179	CKD1179

Parameter	Measurement Range	Kit Part Code (30 tests)	Comparator Disc
Hydrogen Peroxide LR	0 - 1 mg/L H ₂ O ₂	CKH1104	CKD1104
Hydrogen Peroxide HR	0 - 100 mg/L H ₂ O ₂	CKH1105	CKD1105
Iron MR	0 - 10 mg/L Fe	CKH1292	CKD1292
Manganese	0 - 0.03 mg/L Mn	CKH1173	CKD1173
Manganese HR	0 - 5 mg/L Mn	CKH1174	CKD1174
Molybdate HR	0 - 100 mg/L MoO ₄	CKH1175	CKD1175
Nitrate	0 - 15 mg/L N	CKH1163	CKD1163
Nitrite	0 - 0.4 mg/L N	CKH1109	CKD1109
Ozone	0 - 1.0 mg/L O ₃	CKH1056	CKD1056
Phosphate LR	0 - 4 mg/L PO ₄	CKH1177	CKD1177
Phosphate HR	0 - 100 mg/L PO ₄	CKH1114	CKD1114
Silica	0 - 4 mg/L SiO ₂	CKH1181	CKD1181
Sulfide	0 - 0.5 mg/L S	CKH1168	CKD1168
Zinc	0 - 4 mg/L Zn	CKH1148	CKD1148
pH Phenol Red	pH 6.8 - 8.4	CKH1131	CKD1131
Universal pH	pH 4 - 11	CKH1006	CKD1006

Dual-Parameter Kits	Measurement Range	Part Code
Free & Total Chlorine, Phenol Red	0 - 5 mg/L Cl ₂ and pH 6.8 - 8.4	CKH2001
Free & Total Chlorine, Phenol Red	0 - 2 mg/L Cl ₂ and pH 6.8 - 8.4	CKH2002
Free & Total Chlorine, Phenol Red	0 - 1 mg/L Cl ₂ and pH 6.8 - 8.4	CKH2003
Bromine, pH Phenol Red	0 - 8 mg/L Br ₂ and pH 6.8 - 8.4	CKH2005
Free & Total Chlorine, Chlorine HR	0 - 5 and 0 - 250 mg/L Cl ₂	CKH2006
Free & Total Chlorine, Dual Range	0 - 1 and 0 - 5 mg/L Cl ₂	CKH2007
Hydrogen Peroxide, Dual Range	0 - 1 and 0 - 100 mg/L H ₂ O ₂	CKH2104
Copper and pH Phenol Red	0 - 5 mg/L Cu and pH 6.8 - 8.4	CKH2186

Contour Comparator Accessories

Ordering Information	Part Code
Comparator Cuvettes, 10 mL square comparator cuvettes, pack of 5	PT521/5
Dilution Tube, Graduated tube for sample dilution. Graduations at x2, x3, x4, x5, x10	PT512
Cuvette Brush, for cleaning cuvettes, length 120mm	PT663
Crush/Stirring Rods. Pack of 10.	PT502